

**AdipoGen**<sup>®</sup>   
LIFE SCIENCES

## MANUAL

### **FGF-21 (human) ELISA Kit**

[Fibroblast Growth Factor 21 (human) ELISA Kit]

*For research use only. Not for diagnostic use*

Version 1 (May-06-2015)

**Cat. No. AG-45B-5001-KI01**

[www.adipogen.com](http://www.adipogen.com)

## Table of Contents

---

<b>1. Intended Use</b>	<b>3</b>
<b>2. Introduction</b>	<b>3</b>
<b>3. General References</b>	<b>4</b>
<b>4. Assay Principle</b>	<b>5</b>
<b>5. Handling &amp; Storage</b>	<b>5</b>
<b>6. Kit Components</b>	<b>5</b>
<b>7. Materials Required but <i>Not</i> Supplied</b>	<b>6</b>
<b>8. General ELISA Protocol</b>	<b>7</b>
8.1. Preparation and Storage of Reagents	7
8.2. Sample Collection, Storage and Dilution	8
8.3. Assay Procedure (Checklist)	9
<b>9. Calculation of Results</b>	<b>10</b>
<b>10. Typical Data</b>	<b>10</b>
<b>11. Performance Characteristics</b>	<b>11-12</b>
<b>12. Technical Hints and Limitations</b>	<b>13</b>
<b>13. Troubleshooting</b>	<b>14</b>
<b>14. Notes</b>	<b>15</b>

## 1. Intended Use

The FGF-21 (human) ELISA Kit is to be used for the *in vitro* quantitative determination of human FGF-21 in cell culture supernatants, serum and plasma. This ELISA Kit is for research use only.

## 2. Introduction

Fibroblast growth factor 21 (FGF-21) is a novel protein that has been implicated in the regulation of lipid and glucose metabolism under fasting and ketotic conditions (1, 2). In murine models, FGF-21 is predominantly expressed in liver, but it also expressed in brown and white adipose tissues and pancreatic  $\beta$ -cells. FGF-21 is also induced in skeletal, heart, and gastrocnemius muscle under local and systemic metabolic stress (3). FGF-21 acts through the receptor  $\beta$ -Klotho-FGFR (4) and induces expression of Adiponectin that mediates its effects (5).

FGF-21 stimulates glucose uptake in adipocytes. It also protects animals from diet-induced obesity when overexpressed in transgenic mice and lowers blood glucose and triglyceride levels when administered to diabetic rodents. When administered daily for 6 weeks to diabetic rhesus monkeys, FGF-21 caused a dramatic decline in fasting plasma glucose, fructosamine, triglycerides, insulin, and glucagon (6). Furthermore, elevated plasma FGF-21 concentrations in humans appear to be related to the presence of hepatic and peripheral insulin resistance (7). Recently it has been demonstrated that FGF-21 controls circadian function in both brain and peripheral tissues (8).

### 3. General References

- (1) FGF-21 as a novel metabolic regulator: A. Kharitonov, et al.; *J. Clin. Invest.* **115**, 1627 (2005)
- (2) Hepatic fibroblast growth factor 21 is regulated by PPARalpha and is a key mediator of hepatic lipid metabolism in ketotic states: M.K. Badman, et al.; *Cell Metab.* **5**, 426 (2007)
- (3) Identification of a novel FGF, FGF-21, preferentially expressed in the liver: T. Nishimura, et al.; *Biochim. Biophys. Acta* **1492**, 203 (2000)
- (4) Tissue-specific expression of betaKlotho and fibroblast growth factor (FGF) receptor isoforms determines metabolic activity of FGF19 and FGF21: F. Kurosu, et al.; *J. Biol. Chem.* **282**, 26687 (2007)
- (5) An FGF21-adiponectin-ceramide axis controls energy expenditure and insulin action in mice: W.L. Holland, et al.; *Cell Metab.* **17**, 790 (2013)
- (6) The metabolic state of diabetic monkeys is regulated by fibroblast growth factor-21: A. Kharitonov, et al.; *Endocrinology* **148**, 774 (2007)
- (7) Circulating fibroblast growth factor-21 is elevated in impaired glucose tolerance and type 2 diabetes and correlates with muscle and hepatic insulin resistance: A.O. Chavez, et al.; *Diabetes Care* **32**, 1542 (2009)
- (8) FGF21 regulates metabolism and circadian behavior by acting on the nervous system: A.L. Bookout, et al.; *Nature Medicine* **19**; 1147 (2013)

## 4. Assay Principle

This assay is a sandwich Enzyme Linked-Immunosorbent Assay (ELISA) for quantitative determination of human FGF-21 in cell culture supernatants, serum and plasma. A polyclonal antibody specific for FGF-21 has been precoated onto the 96-well microtiter plate. Standards (STD) and samples are pipetted into the wells for binding to the coated antibody. After extensive washing to remove unbound compounds, FGF-21 is recognized by the addition of a biotinylated polyclonal antibody specific for human FGF-21 (DET). After removal of excess biotinylated antibody, streptavidin-peroxidase (STREP-HRP) is added. Following a final washing, peroxidase activity is quantified using the substrate 3,3',5,5'-tetramethylbenzidine (TMB). The intensity of the color reaction is measured at 450 nm after acidification and is directly proportional to the concentration of human FGF-21 in the samples.

## 5. Handling & Storage

- Reagents must be stored at 2-8°C when not in use.
- Plate and reagents should be at room temperature before use.
- Do not expose reagents to temperatures greater than 25°C.

## 6. Kit Components

- 1 vial human FGF-21 Standard (lyophilized) (1920 pg) (STD)
- 1 vial Detection Antibody (100x) (120 µl) (DET)
- 1 vial HRP Labeled Streptavidin (200x) (60 µl) (STREP-HRP)
- 2 bottles Wash Buffer 10X (2 x 25 ml) (Wash Buffer 10X)
- 1 bottle Assay Buffer 5X (20 ml) (Assay Buffer 5X)
- 1 bottle TMB Substrate Solution (12 ml) (TMB)
- 1 bottle Stop Solution (12 ml) (STOP)
- 1 plate coated with FGF-21 Antibody (12 x 8-well strips)
- 2 plate Covers (plastic film)
- 2 silica Gel Minibags

## 7. Materials Required but *Not* Supplied

- Microtiterplate reader at 450nm
- Calibrated precision pipettes. Disposable pipette tips
- Deionized water
- Microtubes or equivalent for preparing dilutions
- Disposable plastic containers for preparing working buffers
- Plate washer: automated or manual
- Glass or plastic tubes for diluting and aliquoting standard

## 8. General ELISA Protocol

### 8.1. Preparation and Storage of Reagents

**NOTE:** Prepare just the appropriate amount of the buffers necessary for the assay.

- **Wash Buffer 10X** has to be diluted with deionized water 1:10 before use (e.g. 50 ml Wash Buffer 10X + 450 ml distilled or deionized water) to obtain Wash Buffer 1X. The 1 x Wash buffer may be stored at 2-8°C for up to one month.
- **Assay Buffer 5X** has to be diluted with deionized water 1:5 before use (e.g. 10 ml Assay Buffer 5X + 40 ml distilled or deionized water) to obtain Assay Buffer 1X. The 1 x Assay buffer may be stored at 2-8°C for up to one month.
- **Detection Antibody (DET) 100x** (after a spin-down) has to be diluted to 1:100 in Assay Buffer 1X (20 µl DET + 1.98 ml Assay Buffer 1X). 100µl of the 1x Detection Antibody is required per well.

**NOTE:** Prepare only as much 1x Detection antibody as needed. The diluted Detection Antibody is not stable and cannot be stored!

- **HRP Labeled Streptavidin (STREP-HRP) 200X** (after a spin-down) has to be diluted to 1:200 with Assay Buffer 1X (50 µl in 9.95 ml of Assay Buffer 1X). 100µl of the 1x STREP-HRP is required per well.

**NOTE:** Prepare only as much 1x STREP-HRP as needed. The diluted STREP-HRP is not stable and cannot be stored!

- **Human FGF-21 Standard (STD)** has to be reconstituted with 1ml of Assay Buffer 1X.
  - This reconstitution produces a stock solution of 1920 pg/ml. Mix the standard to ensure complete reconstitution and allow the standard to sit for a minimum of 15 minutes **at RT**. Mix well prior to making dilutions.

**NOTE:** The reconstituted standard is aliquoted and stored at -20°C!

- Dilute the standard protein concentrate (STD) in Assay Buffer 1X. A seven-point standard curve using 2-fold serial dilutions in Assay Buffer 1X is recommended.
- Suggested standard points are:  
**1920, 960, 480, 240, 120, 60, 30, and 0 pg/ml.**

To obtain	Add	Into
1920 pg/ml	1920 pg/ml stock	-
960 pg/ml	250 µl of FGF-21 (1920 pg/ml )	250 µl of Assay buffer 1X
480 pg/ml	250 µl of FGF-21 (960 pg/ml)	250 µl of Assay buffer 1X
240 pg/ml	250 µl of FGF-21 (480 pg/ml)	250 µl of Assay buffer 1X
120 pg/ml	250 µl of FGF-21 (240 pg/ml)	250 µl of Assay buffer 1X
60 pg/ml	250 µl of FGF-21 (120 pg/ml)	250 µl of Assay buffer 1X
30 pg/ml	250 µl of FGF-21 (60 pg/ml)	250 µl of Assay buffer 1X
0 pg/ml	250 µl of Assay buffer 1X	Empty tube

**NOTE:** Do not store the diluted standard solutions.

## 8.2. Sample Collection, Storage and Dilution

**Cell Culture Supernatants, serum and plasma** have to be diluted in Assay Buffer 1X. Starting dilution of 1/2 is recommended.



### 8.3. Assay Procedure (Checklist)


<input type="checkbox"/>	<p>1. Determine the number of 8-well strips needed for the assay and insert them in the frame for current use. The extra strips are left in the bag with 2 silica gel minibags and stored at 4°C.</p> <p><b>NOTE:</b> Remaining 8-well strips coated with FGF-21 antibody when opened can be stored in the presence of 2 silica gel minibags at 4°C for up to 1 month.</p>
<input type="checkbox"/>	<p>2. Add 100 µl of the different standards into the appropriate wells in duplicate! At the same time, add 100 µl of diluted serum, plasma or cell culture supernatant samples in duplicate to the wells (<b>see 8.1. Preparation and Storage of Reagents and 8.2 Preparation of Samples</b>).</p>
<input type="checkbox"/>	<p>3. Cover the plate with plastic film and incubate for <b>1 h at room temperature</b>.</p>
<input type="checkbox"/>	<p>4. Aspirate the coated wells and add 300 µl of Wash Buffer 1X using a multichannel pipette or auto-washer. Repeat the process for a total of five washes. After the last wash, complete removal of liquid is essential for good performance.</p>
<input type="checkbox"/>	<p>5. Add 100 µl to each well of the 1x Detection Antibody (<b>DET</b>) (<b>see 8.1 Preparation and Storage of Reagents</b>).</p>
<input type="checkbox"/>	<p>6. Cover the plate with plastic film and incubate for <b>1 h at room temperature</b>.</p>
<input type="checkbox"/>	<p>7. Aspirate the coated wells and add 300 µl of Wash Buffer 1X using a multichannel pipette or auto-washer. Repeat the process for a total of five washes. After the last wash, complete removal of liquid is essential for good performance.</p>
<input type="checkbox"/>	<p>8. Add 100 µl to each well of the diluted HRP Labeled Streptavidin (<b>STREP-HRP</b>) (<b>see 8.1. Preparation and Storage of Reagents</b>).</p>
<input type="checkbox"/>	<p>9. Cover the plate with plastic film and incubate for <b>20 min at room temperature</b>.</p>
<input type="checkbox"/>	<p>10. Aspirate the coated wells and add 300 µl of Wash Buffer 1X using a multichannel pipette or auto-washer. Repeat the process for a total of five washes. After the last wash, complete removal of liquid is essential for good performance.</p>
<input type="checkbox"/>	<p>11. Add 100 µl to each well of TMB substrate solution (<b>TMB</b>).</p>
<input type="checkbox"/>	<p>12. Allow the color reaction to develop <b>at room temperature in the dark for 15 minutes</b>. Do not cover the plate.</p>
<input type="checkbox"/>	<p>13. Stop the reaction by adding 100 µl of Stop Solution (<b>STOP</b>). Tap the plate gently to ensure thorough mixing. The substrate reaction yields a blue solution that turns yellow when Stop Solution (<b>STOP</b>) is added.</p>
<input type="checkbox"/>	<p><b>! CAUTION: CORROSIVE SOLUTION !</b></p>
<input type="checkbox"/>	<p>14. Measure the OD at 450 nm in an ELISA reader.</p>

## 9. Calculation of Results

- Average the duplicate readings for each standard, control and sample and subtract the average blank value (obtained with the 0 pg/ml point).
- Generate the standard curve by plotting the average absorbance obtained for each standard concentration on the horizontal (X) axis vs. the corresponding FGF-21 concentration (pg/ml) on the vertical axis (see **10. TYPICAL DATA**).
- Calculate the FGF-21 concentrations of samples by interpolation of the regression curve formula as shown above in a form of a quadratic equation
- If the test sample was diluted, multiply the interpolated value by the dilution factor to calculate the concentration of human FGF-21 in the sample.

## 10. Typical Data

The following data are obtained using the different concentrations of standard as described in this protocol:



Human FGF-21 (pg/ml)	Absorbance (450nm)
0	0.094
30	0.127
60	0.158
120	0.224
240	0.362
480	0.626
960	1.141
1920	2.14

**Figure: Standard curve**

## 11. Performance Characteristics

### A. Sensitivity (Limit of detection):

The lowest level of human FGF21 that can be detected by this assay is 7 pg/ml. **NOTE:** *The Limit of detection was measured by adding three standard deviations to the mean value of 50 zero standard.*

**B. Assay range:** 30 pg/ml – 1920 pg/ml

### C. Specificity:

This ELISA is specific for the measurement of natural and recombinant human FGF-21. It does not cross-react with human FGF-19 and FGF-23.

### D. Intra-assay precision:

Two samples of known concentrations of human FGF-21 were assayed in replicates 12 times to test precision within an assay.

Samples	Means (ng/ml)	SD	CV (%)	n
A1	93.3	3.72	4	12
A2	547	27.6	5	12

### E. Inter-assay precision:

Two samples of known concentrations of human FGF-21 were assayed in 10 separate assays to test precision between assays.

Samples	Means (ng/ml)	SD	CV (%)	n
B1	273.5	27.9	10.2	10
B2	335	11.8	3.5	10

**F. Linearity:**

To assess the linearity of the assay, samples containing and/or spiked with high concentrations of human FGF-21 were serially diluted with the 1X Assay buffer to produce samples with values within the dynamic range of the assay.

## Sample 1

Dilution	Measured (pg/ml)	Expected (pg/ml)	Recovery (%)
None	448	448	100
1/2	245	224	109
1/4	123	112	109
1/8	63.6	56	113

## Sample 2

Dilution	Measured (pg/ml)	Expected (pg/ml)	Recovery (%)
None	318	318	100
1/2	170	159	106
1/4	84.3	79.5	106
1/8	43.1	39.75	108

**G. Expected values:**

Human FGF-21 levels range in plasma or serum from **non-detectable to >1000pg /ml**.


## 12. Technical Hints and Limitations

- It is recommended that all standards, controls and samples be run in duplicate.
- Do not combine leftover reagents with those reserved for additional wells.
- Reagents from the kit with a volume less than 100µl should be centrifuged.
- Residual wash liquid should be drained from the wells after last wash by tapping the plate on absorbent paper.
- Crystals could appear in the 10X solution due to high salt concentration in the stock solutions. Crystals are readily dissolved at room temperature or at 37°C before dilution of the buffer solutions.
- Once reagents have been added to the 8-well strips, DO NOT let the strips DRY at any time during the assay.
- Keep TMB Solution protected from light.
- The Stop Solution (STOP) consists of sulfuric acid. Although diluted, the Stop Solution should be handled with gloves, eye protection and protective clothing.

## 13. Troubleshooting

PROBLEM	POSSIBLE CAUSES	SOLUTIONS
No signal or weak signal	Omission of key reagent	Check that all reagents have been added in the correct order.
	Washes too stringent	Use an automated plate washer if possible.
	Incubation times inadequate	Incubation times should be followed as indicated in the manual.
	Plate reader settings not optimal	Verify the wavelength and filter setting in the plate reader.
	Incorrect assay temperature	Use recommended incubation temperature. Bring substrates to room temperature before use.
High background	Concentration of STREP-HRP too high	Use recommended dilution factor.
	Inadequate washing	Ensure all wells are filling wash buffer and are aspirated completely.
Poor standard curve	Wells not completely aspirated	Completely aspirate wells between steps.
	Reagents poorly mixed	Be sure that reagents are thoroughly mixed.
Unexpected results	Omission of reagents	Be sure that reagents were prepared correctly and added in the correct order.
	Dilution error	Check pipetting technique and double-check calculations.

## 14. Notes



**Adipogen Life Sciences**  
Schützenstrasse 12  
CH-1410 Liestal  
Switzerland  
TEL: +41-61-926-60-40  
FAX: +41-61-926-60-49  
Email: [info@adipogen.com](mailto:info@adipogen.com)

**AdipoGen<sup>®</sup>**   

---

**LIFE SCIENCES**